

**Department of Optoelectronics, University of Kerala, Kariavattom,
Thiruvananthapuram, Kerala, India – 695581, Ph: 91 471 2308167**

Quotation Notice

OPTO/P-20/2022-23

Quotations are invited for the following item in the Department of Optoelectronics, University of Kerala, Kariavattom, Thiruvananthapuram

Analytical weighing balance

Preferable make: Denver, Sartorius, Mettler, Citizen

Specifications

- Readability : 0.01 mg
- Capacity : 210gm
- Repeatability : 0.02mg
- Minimum capacity: 0.001g
- Response time : $\leq 12/3$
- Pan size : $\geq \varnothing 80\text{mm}$
- Internal calibration

The price may clearly indicate installation, applicable tax, packing, forwarding and handling charges separately. The firms wish to supply the item are requested to submit the quotation in sealed cover to the undersigned not later than 3:30pm on 09.11.2022. Quotations will be opened at 10:30am on next working day (10.11.2022) in the presence of the suppliers or their authorized agents.

26.10.2022

Sd/-
Professor & Head