

Quotations are invited for the following instrument from the manufactures/authorised suppliers

Instrument :COD MEASUREMENT EQUIPMENT

(as per the below given details and specifications)

Description : Thermoreactors for COD Analysis - 25-position thermoreactors for Chemical Oxygen Demand (COD) with settable temperature (70, 100, 120, 150, 160 °C) and operating time (30, 60, 120 minutes or continuous).

Technical Specifications	
Construction material:	Epoxy painted metal structure
Models:	ECO 25 25 pos. (Ø16 mm)
Set temperature:	display visualization
Countdown:	display visualization
End of the cycle:	acoustic signal with automatic switch-off
PERFORMANCE	
Selectable working temperatures:	70, 100, 120, 150 and 160
Selectable working times:	30, 60, 120 or continuous
HEATING BLOCK TECHNICAL DATA	
Temperature stability:	± 0.5 °C
Temperature homogeneity:	± 0.5 °C
Temperature precision:	± 1 °C
Overtemperature safety	
SIGNALS	
Reaching of set temperature:	acoustic and visual signal
Countdown:	visual signal
End of cycle:	acoustic and visual signal
Probe interruption:	acoustic and visual signal
Broken probe:	acoustic and visual signal
Exceed temperature range:	acoustic and visual signal

TERMS & CONDITION

Delivery : At Kariavattom Campus
Rate : Inclusive of tax
Payment : by demand draft or Cheque

Interested parties may submit their quotations on or before 4pm of 25th October 2017 to the following address:

Co-ordinator
International and Inter University Centre for Natural Resources Management
University of Kerala
Kariavattom campus
Trivandrum 695 581