

**Department of Nanoscience and Nanotechnology,
University of Kerala,
Kariavattom, Thiruvananthapuram-695581
Ph.No. 0471 2308863**

Re-Tender Notice

No. KU/DNST/QN- 02/2017-18

Quotations are invited for the supply of Rectangular Vacuum Oven with the following specifications. Re-Tendering was necessitated due to the lack of clarity in the specification of the Vacuum pump required.

Name of equipment	Technical Specifications
Rectangular Vacuum Oven	Temperature range- 50°C to 250°C. The inner chamber is made out of SS and outer of MS sheet duly powder coated glass wool insulation is provided all around the chamber to minimize heat loss due to radiation. The door and its frame is made of heavy gauge extruded Aluminum having a toughened glass window to view the samples placed inside. The temperature is controlled and indicated by a micro processor based digital temperature controller (LED) vacuum control valve with vacuum gauge. Maximum vacuum: 760mm of Hg. Chamber size - 450 mm×450 mm× 450 mm Warranty – 1 year Vacuum pump – Minimum ¼ hp or above.

GENERAL INSTRUCTIONS:-

1. Incomplete & conditional quotations and quotations received after the due date will be summarily rejected without assigning any reasons thereof.
2. The price should be inclusive of all taxes, duties, transportation, installation etc. Nothing extra will be paid in addition to the quoted rate.
3. Payment Terms: 100% payment after supply and successful demonstration of the working of the equipment.

4. Validity of quotation: Quotation submitted shall remain valid at least for 90 days from the date of opening the quotation.

5. Delivery and installation: Proposed delivery schedule should be mentioned clearly. Delivery and installation should be made at the Department of Nanoscience and Nanotechnology, University of Kerala, Kariavattom campus, Trivandrum- 695581, without any extra cost.

6. Service facility: Supplier should mention their details of service setup and manpower in Trivandrum who are responsible for after sales support.

7. The model number, make, and a printed literature of the product shall be submitted positively.

8. Vendors should provide a list of institutes showing installations of such instruments in India with contact numbers.

Sealed quotations should reach the “Head-in-charge, Department of Nanoscience and Nanotechnology, University of Kerala, Kariavattom, Thiruvananthapuram-695 581” on or before 03-07-2017, 3:00 pm.

S/d

19-06-2017

Head-in-charge