

**Department of Optoelectronics, University of Kerala, Kariavattom,
Thiruvananthapuram, Kerala, India – 695581, Ph: 91 471 2308167**

Quotation Notification

OPTO/P-06/2017-18

Quotations are invited for the following item in the Department of Optoelectronics,
University of Kerala, Kariavattom, Thiruvananthapuram


Item: Centrifuge

Specifications:

Microprocessor controlled
Imbalance detector with cut-off
Safety lid interlock to prevent cover opening during centrifugation
Capacity required: 8 x 25 ml
Maximum Speed required: 14500 rpm
Sample Containers are to be provided

The price may clearly indicate applicable tax, packing, forwarding and handling charges (if any) separately. The firms who wish to supply the item are requested to submit the quotation in sealed cover to the undersigned not later than 4:30pm on 18.01.2018. Quotations will be opened at 10:30 am on next working day (19.01.2018) in the presence of the suppliers or their authorized agents.

30.12.2017



A handwritten signature in black ink, consisting of several horizontal strokes and a vertical stroke, positioned above the printed name.

Associate Professor & Head
Associate Professor & Head
Department of Optoelectronics
University of Kerala, Kariavattom P.O
Thiruvananthapuram-695581