

**Department of Optoelectronics, University of Kerala, Kariavattom,
Thiruvananthapuram, Kerala, India – 695581, Ph: 91 471 2308167**

Quotation Notice

OPTO/P-2/2016-17

Quotations are invited for the following item in the Department of Optoelectronics, University of Kerala, Kariavattom, Thiruvananthapuram

Analytical weighing balance

Preferable make: Denver, Sartorius, Mettler, Citizen

Specifications

- Readability : 0.01 mg
- Capacity : 210gm
- Repeatability : 0.02mg
- Minimum capacity: 0.001g
- Response time : $\leq 12/3$
- Pan size : $\geq \varnothing 80\text{mm}$
- Internal calibration

The firms who wish to supply the items are requested to submit the quotation in sealed cover to the undersigned on or before 25.05.2016.

11.05.2016




Professor & Head
PROFESSOR & HEAD
Department of Optoelectronics
University of Kerala, Kariavattom
Thiruvananthapuram - 695 581