University of Kerala

Department of Physics

Kariavattom campus, Thiruvananthapuam -695581

KERALA

No.Phy/209/2016-17

27-10-2016

Quotation Notice

Quotations are invited for purchasing branded laboratory centrifuge for research purpose with following specification by the Department of Physics, University of Kerala.

| No | Item | Description/Remark |
|----|-----------------------|---|
| 1 | Laboratory centrifuge | Laboratory centrifuge having following specifications 1. Type of product: Laboratory centrifuge 2. Capacity: 400 ml or more 3. No of tubes: preferably 4 or 6 4. Max speed: 7000 rpm or more 5. Digital timer: 0 – 60 minutes* 6. supply: 220-240V, 50 HZ 7. Dimension: 380 mm x 470 mm x 300mm. 8. Warranty not less than three years. |
| | | The centrifuge and accessories should be delivered and installed at the Department of Physics University of Kerala. |

Quotes should be inclusive of the cost, freight, taxes, etc. Terms and conditions of warranty should be clearly mentioned. Please quote the lowest prize for the item. The quotation should reach "Professor and Head, Department of Physics, University of Kearla, Kariavttom Campus, Thiruvananthapuram, Pin-695 581, Kerala, India" on or before 14-11-2016. 3PM.

HIRUVANANTHAPU

Sd/-

Dr. Subodh G.
Assistant professor and Head
Department of Physics
University Of Kerala
Karivattom, Thiruvananthapuram.

Head
Department of Physics
University of Kerala
Kariavattom-695 581